

JM Test systems

Test Equipment Made Easy

CALIBRATION-SALES-RENTAL-ELECTRICAL SAFETY

1020 N. Texas Avenue Odessa, TX 79761
Phone: 866-631-9091 Fax: 432-580-5030

Baton Rouge, LA - Alexandria, LA - Clute, TX - Odessa, TX - Mattoon, IL
Port Arthur, TX - San Angelo, TX - Berthold, ND - Mount Braddock, PA - Mathiston, MS

This work was performed by our Odessa location, Lab division.

CERTIFICATE OF CALIBRATION

Manuf. ASHCROFT
Model 0-400 PSI

Mfg Serial 8486
Cust ID# N/A

Cust Serial N/A

Desc. PRESSURE GAUGE

PO# W/OPO
Submitted By
Gage Western LLC
7907 W Industrial
Midland, TX 79706

JM TEST SYSTEMS certifies this instrument has been repaired, if necessary, and calibrated to published manufacturer specifications using standards traceable to NIST, to consensus standards, to ratio methods, or to accepted values of natural physical constants. This certificate applies only to the item listed above. JM TEST SYSTEMS' quality control system meets the requirements of ANSI/NCSL Z540-1, ISO 9001, ISO 10012 and ISO 17025.

Condition In: In Tolerance
Condition Out: No Adjustments
Action: Cleaned and certified.
Procedure: JM011 Procedure Rev. 07/27/2022

The following standards were used in the calibration:

Manufacturer/Model	Description	Std #	Date Due
CRYSTAL ENGINEERING / 1KPSIXP2I	DIGITAL TEST GAUGE	3652	11/30/2023

This report shall not be reproduced, except in full, without the written approval of JM TEST SYSTEMS, LLC.

Certificate/traceability number: 12898.00-464024-001


Date of Report: 09/27/2023

WO#: 464024-001

Calibration Date: 09/27/2023

Due Recalibration Date: 09/27/2024

Lab Environment: TEMP: 24°C
HUMIDITY: 34%RH


Technician Kasey W Johnson



JM Test Systems, Inc.

Manufacturer: ASHCROFT

Description: PRESSURE GAUGE

Mfr S/N: 8486

Calibration Data Sheet
Model: 0-400 PSI

Work Order: 464024-001

Cust ID: N/A

Cust S/N: N/A

Function / Range		Out	Measured Value		Acceptance Limits	
Test Description	Nominal Value	*	AS FOUND	AS LEFT	LOW	HIGH
TEST RESULT						
I/A/W JM011 Procedure:						
Pressure Gauge: 0-400 psi						
0-400 psi Range:						
Upscale Accuracy:						
0 psi	0 psi		0	N/A	-2 psi	2 psi
100 psi	100 psi		100	N/A	98 psi	102 psi
200 psi	200 psi		200	N/A	198 psi	202 psi
300 psi	300 psi		300	N/A	298 psi	302 psi
400 psi	400 psi		400	N/A	398 psi	402 psi
Downscale Accuracy:						
300 psi	300 psi		300	N/A	298 psi	302 psi
200 psi	200 psi		200	N/A	198 psi	202 psi
100 psi	100 psi		100	N/A	98 psi	102 psi
0 psi	0 psi		0	N/A	-2 psi	2 psi