



**W.L. WALKER CO. INC**  
Established 1926

# Thermometer Calibration Certificate

Sold To:

GAGE WESTERN

Standard A ☐

Serial #

881693

MIDLAND, TX

Standard B ☐

Range

30

To

220

Divisions

0.2

Catalog #

56218WC

Fahrenheit ☒

Celsius ☐

Manufacturer

Kessler

Type

N/A

Immersion

3 "

Thermometer Readings

32.4 F

Corrections

-0.4 F

122.3 F

-0.3 F

212.4 F

-0.4 F

F

F

F

F

F

F

W. L. Walker certifies the customer thermometer identified above was calibrated with thermometer standards traceable to the National Institute of Standards and Technology (NIST). Primary and transfer standards are maintained as per NIST.

A: Reference Device: Digital Readout:  
Model 1529-R SN A6C413

PRT: Model 5627A, SN 866120

PRT: Model 5627A, SN 866121

PRT: Model 5627A, SN 782477

B: Reference Device: Digital Readout:  
Model 1529-R SN B2C802

PRT: Model 5627A, SN 887388

PRT: Model 5627A, SN 843100

PRT: Model 5627A, SN 887386

**NOTES: The thermometer was tested in a closed-top, electrically heated liquid bath, and was immersed as stated above. The temperature of the tested room ("ambient" temperature) was about 25 Degrees C (77 F). Use of the thermometer at other immersion depths or in greatly different ambient temperatures than stated above will cause appreciable errors.**

NIST Special Publication 1088 January, 2009



Technician

CHS

Date

10/31/2023

Order#

0049763